



Certificate of Compliance

Certificate Number: LR 91712-12

Revision:

Date Issued: February 17, 1995

Issued To: GREYLINE INSTRUMENTS INC.
800 CAMPBELL ST.
CORNWALL, ONT. K6H 6L7

The products listed below are eligible to bear the CSA Mark.

Issued By: E. Chow
Toronto, ON Canada

Signature

CLASS 2258 03 - PROCESS CONTROL EQUIPMENT - Intrinsically Safe and
Non-Incendive Systems - For Hazardous Locations

PRODUCTS

Class I, Groups C and D; Class II, Groups E, F and G; Class III; Type 4:

- Ultrasonic Transducer, Model SE rated 10V max, 50mA max; intrinsically safe when connected to CSA Certified Safety Barrier, rated 9.6V max and 27 ohms min.