

Declaration of Conformity

We,

Greyline Instruments Inc.
16456 Sixsmith Drive, Long Sault
Ontario, Canada, K0C 1P0

declare under our sole responsibility that the product

Portable Level-Velocity Logger Model: Stingray

to which this declaration relates conforms with the following European, harmonized and published standards at date of this declaration:

Environment: Industrial

EN 50081-2:1993 Generic emission standard, from which:

EN 55011:1998	Emission – Class A
EN 61000-3.2:1995	Harmonic current emissions

EN 50082-2:1995 Generic immunity standard, from which:

EN 61000-4-2:1995	Electrostatic discharge (ESD) immunity
EN 61000-4-3:1996	Radiated Electro-Magnetic field immunity
EN 50204:1995	Digital radio telephones immunity
EN 61000-4-4:1995	Electrical fast transient (EFT) immunity
EN 61000-4.5:1995	Surge transient immunity
EN 61000-4-6:1996	Conducted Radio-frequency disturbances immunity

following the provisions of the Directives (if applicable):

EMC-directive : 89/336/EEC

These conclusions are based on test reports:

Report Number 96540-KRQ/EMC 99-4616 and 2006416-KRQ/EMC 00-4619 issued by:

KEMA Registered Quality B.V.
Arnhem, The Netherlands

And

Greyline Engineering Report, entitled "Stingray CE Requirements Review",
dated 11 March, 2002

Long Sault, Ontario Thursday March 18, 2004

W.F. Robson, President