

# Declaration of Conformity

We,

Greyline Instruments Inc.  
16456 Sixsmith Drive,  
Long Sault, Ontario, K0C 1P0

---

Declare under our sole responsibility that the product

Area Velocity Flow Meter, Model: AVFM 5.0

---

To which this declaration relates conforms with the following European, harmonized and published standards at date of this declaration:

**EN 61326-1:2006** *Electrical equipment for measurement, control and laboratory use- EMC requirements-Part 1: General Requirements*

CISPR 11  
IEC 61000-3-2  
IEC 61000-3-3  
IEC 61000-3-11  
IEC 61000-3-12  
IEC 61000-4-2  
IEC 61000-4-3  
IEC 61000-4-4  
IEC 61000-4-5  
IEC 61000-4-6  
IEC 61000-4-11  
IEC 61010-1

En61326-2-3:2009 Electrical equipment for measurement control and laboratory use EMC requirements. Part 2-3: Particular requirements - Test configuration and performance criteria for transducers with integrated or remote signal conditioning.

---

Long Sault, Ontario, 18<sup>th</sup> February, 2011

---

W.F. Robson, President

